

EC type examination certificate:	EX5 12 02 45670 008
Ex-identification:	⊕ II 2 G Ex ib IIB T3 -10 °C ≤ Ta ≤ +50 °C
Power supply:	4.5VDC
Current input:	max. 60mA
Minimum conductivity:	0,8mS/cm
Measuring voltage:	3,3Veff
Measuring current:	max. 1,3mA
Measuring frequency:	10Hz
Working temperature:	-10 °C bis +50 °C
Storage temperature:	-25 °C bis +70 °C (without batteries)
Housing:	Vestamide/black (antistatic)
LxWxH:	180 x 85/62 x 30/25.5 mm
Weight:	390 g
Degree of protection:	IP 54
Memory:	Owner: 1 entry User: 50 entries Locations: 500 entries Measurements: 10,000 entries (circulating)
Sealed keypad:	21 keys (shielded)
Battery compartment for:	3 batteries size R6
Permitted batteries:	AA 1.5 V size R6 LR6 No. 4106 AM3 MN 1500 VARTA Microbattery GmbH
Graphic display 128x64 and acryl glass cover (metalized)	
Connectors for the immersion probe and grounding clamp on face side of enclosure	
Electromagnetic compatibility (to IEC 801)	
CE mark	

Pf
Pfautler
Defining the standard

Pfautler GmbH
P.O. Box 1780 D-68721 Schwetzingen
Pfautlerstraße D-68723 Schwetzingen
Phone +49 6202 85-233
Telefax +49 6202 85-273
E-mail info@pfautler-instrumentation.com

www.pfautler-instrumentation.com